

**Test Tube Shakers Reax top / control****Heidolph****Test Tube Shakers Reax top / control**

Ideal for mixing samples in Eppendorf tubes, vials and similar vessels of different diameters at extremely high speed rates. The strong 5 mm vibration orbit yields the best results even with high viscosity media or solids out of solution and quickly gives a smooth and even dispersion. Short time and continuous mode.

**Standard model Reax top:**

-Features an analog dial speed control, no settings of specified rpm numbers

**Precise model Reax control additionally with:**

- Features analog dial speed control with accurate rpm numbers setting
- Electronic speed control provides constant rpm output even at very low speed and under changing loads
- Precise speed control with settings of specified rpm numbers

**Scope of delivery:** Test tube shaker incl. attachment diam. 20 mm. Optional attachment for vessels up to 50 mm available.

**Specifications**

**Type of movement:** orbital

**Orbit:** 5 mm

**Speed range:**

**Power consumption:** 0 ... 2500 rpm

51 W

**Permissible ambient temperature:** 5 ... 31 °C

**Permissible relative moisture:** 80 %

**Dimensions (W x D x H):** 134 x 172 x 105 mm

**Weight:** 2.8 kg

**Power supply:** 230 V, 50 Hz

**IP code:** IP 22



Code	Description	Packaging
LLG09730019	Accessories for Test tube shakers Reax top / control, Attachment, large, for tubes or flasks up to 50 ml	1 pz.
LLG09730040	Test tubes shakers REAX top,with small attachment plate,shaking frequency: 0 - 2400 rpm	1 pz.
LLG09730041	Test tube shaker REAX-Top-UK with small attachment with UK plug	1 pz.
LLG09730042	Test tubes shakers REAX control,with small attachment plate,shaking frequency: 0 - 2400 rpm with electronic speed controller	1 pz.



**CARLO ERBA Reagents S.A.S.**  
 Chaussée du Vexin – 27106 Val  
 de Reuil cedex  
 N° TVA: FR 63391048824  
 Tél : +33 (0)2 32 09 20 00  
[www.carloerbareagents.com](http://www.carloerbareagents.com)



**CARLO ERBA**  
*Reagents operates with  
 a Certified Quality  
 Management System*

