


**Graduated pipettes, Class AS, amber graduations, type 2****BRAND****Graduated pipettes, Class AS, amber graduations, type 2**

Blaubrand®, DE-M marking, AR-glas®. DIN EN ISO 835. Type 2 (zero at top), total delivery. Calibrated to deliver (TD, EX). Waiting time 5 seconds. Marks and inscriptions in high contrast blue enamel. All Blaubrand® graduated pipettes are supplied with a batch number and an accompanying batch certificate in the original packaging. Also available with individual works, USP, or DAkkS (former DKD) calibration certificate on request. Supplementary to the DIN range. The batch and single certificates are also available to download from the supplier's website.





CARLO ERBA Reagents S.A.S.
Chaussée du Vexin – 27106 Val
de Reuil cedex
N° TVA: FR 63391048824
Tél : +33 (0)2 32 09 20 00
www.carloerbareagents.com



CARLO ERBA
*Reagents operates with
a Certified Quality
Management System*

