

**Caps for TrayCell****Hellma Analytics****Caps for TrayCell**

Traditional spectrophotometrical analysis of nucleic acids is carried out in 10mm cuvettes. Modern drop analysis uses light paths of 1mm and 0.2mm and it is thus not necessary to dilute the sample. In applications with very high or very low DNA concentrations these light paths are often no longer sufficient. When using the TrayCell from Hellma Analytics simply adjust the light path by exchanging the cap. Cap with integrated mirror to change the light path.



Code	Description	Packaging
LLG09145706	Caps for TrayCell, Path length 0,1 mm	1 pz.
LLG09145705	Caps for TrayCell, Path length 2 mm	1 pz.



CARLO ERBA Reagents S.A.S.
 Chaussée du Vexin – 27106 Val
 de Reuil cedex
 N° TVA: FR 63391048824
 Tél : +33 (0)2 32 09 20 00
www.carloerbareagents.com



CARLO ERBA
*Reagents operates with
 a Certified Quality
 Management System*

