



Carlo Erba Reagents S.A.S.  
 Parc d'Affaires des Portes  
 Chaussée du Vexin - BP 616  
 27106 Val de Reuil Cedex  
 Tél. :02.32.09.20.00  
 Fax 02.32.09.20.20

Z.I DE VALDONNE - BP 4  
 13124 PEYPIN  
 Tél. :04.42.32.41.41  
 Fax 04.42.72.41.62

Carlo Erba Reagents S.r.l.  
 Via R. Merendi, 22  
 20010 Cornaredo (Mi)  
 Tel. 02 93 991 90 Fax 02 93 991 001

ISO 9001: 2015

# Technical Specification

**PRODUCT : BUFFER SOLUTION pH 1 RPE - For analysis**

**CODE : 486210**

| TEST                       | U.M.     | SPECIFICATION |
|----------------------------|----------|---------------|
| Clear, colourless solution | -        | Conform       |
| pH                         | unite pH | 0.98 - 1.02   |
| Temperature of measurement | °C       | 19 - 21       |

Accuracy : +/- 0.02 pH units - Standardized against NIST reference materials.

It is under the customer responsibility to make sure the purchased product is suitable for his use and/or application.  
 CARLO ERBA Reagents would not be guilty of any misuse or mishandling of any of its products, occurring potential damages or user hurts in case of in

Date : 07/10/2025

**QUALITY CONTROL RESPONSIBLE**  
**J. RICHARD (VDR)**