



Carlo Erba Reagents S.A.S.
 Parc d'Affaires des Portes
 Chaussée du Vexin - BP 616
 27106 Val de Reuil Cedex
 Tél. : 02.32.09.20.00
 Fax 02.32.09.20.20

Z.I DE VALDONNE - BP 4
 13124 PEYPIN
 Tél. : 04.42.32.41.41
 Fax 04.42.72.41.62

Carlo Erba Reagents S.r.l.
 Via R. Merendi, 22
 20010 Cornaredo (Mi)
 Tel. 02 93 991 90 Fax 02 93 991 001

ISO 9001: 2015

Technical Specification

PRODUCT : BUFFER SOLUTION pH 4 RPE - For analysis

CODE : 486270

TEST	U.M.	SPECIFICATION
Clear, colourless solution	-	Conform
pH	unite pH	3.98 - 4.02
Temperature of measurement	°C	19 - 21

Accuracy : +/- 0.02 pH units - Standardized against NIST reference materials.

It is under the customer responsibility to make sure the purchased product is suitable for his use and/or application.
 CARLO ERBA Reagents would not be guilty of any misuse or mishandling of any of its products, occurring potential damages or user hurts in case of in

Date : 07/10/2025

QUALITY CONTROL RESPONSIBLE
J. RICHARD (VDR)