

Technical Specification

PRODUCT : Standard solution 147 µS/cm RS For conductivity

CODE : 575060

TEST	U.M.	SPECIFICATION
Description	-	Clear colourless liquid
Identification	-	Positive
Conductivity at 25°C	µS/cm	145.5 - 148.5

Traceable to Standard NIST N° 999a
 Confidence interval: ± 0.1 µS/cm

It is under the customer responsibility to make sure the purchased product is suitable for his use and/or application.
 CARLO ERBA Reagents would not be guilty of any misuse or mishandling of any of its products, occurring potential damages or user hurts in case of in