

Technical Specification

PRODUCT : BUFFER SOLUTION pH 9 +/-0.02 BLUE

CODE : PS0427

TEST	U.M.	SPECIFICATION
pH	unite pH	8.98 - 9.02
Temperature of measurement	°C	19 - 21

Accuracy : +/- 0.02 pH units - Standardized against NIST reference materials.

It is under the customer responsibility to make sure the purchased product is suitable for his use and/or application.
CARLO ERBA Reagents would not be guilty of any misuse or mishandling of any of its products, occurring potential damages or user hurts in case of in