



Carlo Erba Reagents S.A.S.
Parc d'Affaires des Portes
Chaussée du Vexin - BP 616
27106 Val de Reuil Cedex
Tél. : 02.32.09.20.00
Fax 02.32.09.20.20

Z.I DE VALDONNE - BP 4
13124 PEYPIN
Tél. : 04.42.32.41.41
Fax 04.42.72.41.62

Carlo Erba Reagents S.r.l.
Via R. Merendi, 22
20010 Cornaredo (Mi)
Tel. 02 93 991 90 Fax 02 93 991 001

ISO 9001: 2015

Technical Specification

PRODUCT : NITRIC ACID 2 N (2 mol/L) RPE - For analysis

CODE : 408180000

TEST	U.M.	SPECIFICATION
Description	-	Clear colourless liquid
Assay (potentiometry)	N	1.998 - 2.002

This volumetric solution is checked by Tris(hydroxyméthyl)-aminomethan volumetric standard, traceable to the SRM 723e of NIST (National Institute of Standards and Technology)

It is under the customer responsibility to make sure the purchased product is suitable for his use and/or application.
CARLO ERBA Reagents would not be guilty of any misuse or mishandling of any of its products, occurring potential damages or user hurts in case of in

Date : 30/04/2024

QUALITY CONTROL RESPONSIBLE
B. COULANGE (VDR)