

Technical Specification

PRODUCT : SODA LIME RS - For anesthesia and basal metabolism test**CODE : 432870**

TEST	U.M.	SPECIFICATION
Description	-	White granules
Water content	%	12 - 19
Total Alkali Metal	%	<= 4
CO2 activity	%	>= 19.0
Friability	%	~ 95.0
Moisture absorbance	%	<= 7.5
Diameter between 2 and 5 mm	%	>= 78.0
Colour change from white to violet		

It is under the customer responsibility to make sure the purchased product is suitable for his use and/or application.
CARLO ERBA Reagents would not be guilty of any misuse or mishandling of any of its products, occurring potential damages or user hurts in case of in