



Carlo Erba Reagents S.A.S.  
 Parc d'Affaires des Portes  
 Chaussée du Vexin - BP 616  
 27106 Val de Reuil Cedex  
 Tél. : 02.32.09.20.00  
 Fax 02.32.09.20.20

Z.I DE VALDONNE - BP 4  
 13124 PEYPIN  
 Tél. : 04.42.32.41.41  
 Fax 04.42.72.41.62

Carlo Erba Reagents S.r.l.  
 Via R. Merendi, 22  
 20010 Cornaredo (Mi)  
 Tel. 02 93 991 90 Fax 02 93 991 001

ISO 9001: 2015

# Technical Specification

**PRODUCT : Buffer pH 7.20 Weise RS - For analysis**

**CODE : 486410**

TEST	U.M.	SPECIFICATION
Description	-	Clear colourless liquid
pH at 20°C	-	7.15 - 7.25

It is under the customer responsibility to make sure the purchased product is suitable for his use and/or application.  
 CARLO ERBA Reagents would not be guilty of any misuse or mishandling of any of its products, occurring potential damages or user hurts in case of in

Date : 28/05/2024

**QUALITY CONTROL RESPONSIBLE**  
**B. COULANGE (VDR)**