



Carlo Erba Reagents S.A.S.
Parc d'Affaires des Portes
Chaussée du Vexin - BP 616
27106 Val de Reuil Cedex
Tél. : 02.32.09.20.00
Fax 02.32.09.20.20

Z.I DE VALDONNE - BP 4
13124 PEYPIN
Tél. : 04.42.32.41.41
Fax 04.42.72.41.62

Carlo Erba Reagents S.r.l.
Via R. Merendi, 22
20010 Cornaredo (Mi)
Tel. 02 93 991 90 Fax 02 93 991 001

ISO 9001: 2015

Technical Specification

PRODUCT : BUFFER SOLUTION pH 12 RPE - For analysis

CODE : 486690

| TEST | U.M. | SPECIFICATION |
|----------------------------|----------|---------------|
| Clear, colourless solution | - | Conform |
| pH | unite pH | 11.95 - 12.05 |
| Temperature of measurement | °C | 19 - 21 |

Accuracy : +/- 0.05 pH units - Standardized against NIST reference materials.

It is under the customer responsibility to make sure the purchased product is suitable for his use and/or application.
CARLO ERBA Reagents would not be guilty of any misuse or mishandling of any of its products, occurring potential damages or user hurts in case of in

Date : 29/11/2022

QUALITY CONTROL RESPONSIBLE
B. COULANGE (VDR)