

ISO 9001: 2015

Carlo Erba Reagents S.A.S. Parc d'Affaires des Portes Chaussée du Vexin - BP 616 27106 Val de Reuil Cedex Tél.:02.32.09.20.00 Fax 02.32.09.20.20 Z.I DE VALDONNE - BP 4 13124 PEYPIN Tél.:04.42.32.41.41 Fax 04.42.72.41.62 Carlo Erba Reagents S.r.l. Via R. Merendi, 22 20010 Cornaredo (Mi) Tel. 02 93 991 90 Fax 02 93 991 001

Technical Specification

PRODUCT : BUFFER SOLUTION pH 4 RED RPE - For analysis

CODE : 486760

TEST	U.M.	SPECIFICATION
Appearance	-	Pink, Red clear solution
pH	unite pH	3.98 - 4.02
Temperature of measurement	C°	19 - 21

Accuracy : +/- 0.02 pH units - Standardized against NIST reference materials. Shake before use. A deposit can appear during storage. It has no influence on use.

It is under the customer responsibility to make sure the purchased product is suitable for his use and/or application. CARLO ERBA Reagents would not be guilty of any misuse or mishandling of any of its products, occurring potential damages or user hurts in case of in

QUALITY CONTROL RESPONSIBLE B. COULANGE (VDR)